

Listing and evaluation of the results

51396: Dr. Endre Mikus
LabMagister Training and Science Ltd.

Survey of 25 May 2018



Adviser:

Dr. med. Christoph Niederau
MVZ Labor Dortmund Leopoldstraße GbR
Leopoldstraße 10
44147 Dortmund

Tel.: +49 231 860 27 0
Fax: +49 231 860 27 2810
Mail: christoph.niederau@labor-dortmund.de

296

Vitamins 03 - Vitamin C

Analyte	Sam- ple	Unit	Your value	Target value	TV- Type	Lower limit	Upper limit	Deviation	Meets criteria	
Ascorbic acid (Vit. C)	31	mg/l	15.6	17.9	SV	11.5	24.3	-12.8%	+	
	32		9.20	8.57	SV	5.48	11.7	7.4%	+	

Analyte	Sam- ple	Your unit	Stated value	Conversion factor	Method	Manu- facturer	Device
Ascorbic acid (Vit. C)	31	mg/l	15.6	1.00	9999	SD	ZY99
	32		9.20	1.00			



Ascorbic acid (Vit. C) (mg/l, N = 57)										
Collective	Sam- ple	Target value	Review area			Participants collective			Rate (%)	
						AVG	CV	Num.	Sam.	total
all methods	31	17.9	11.5	-	24.3	17.9	10.8	57	96.5	93.0
	32	8.57	5.48	-	11.7	8.57	11.0	57	94.7	

Rate of success: 93%

