

Listing and evaluation of the results

51396: Dr. Endre Mikus
LabMagister Training and Science Ltd.

Survey of 29 May 2019



Adviser:

Dr. med. Christoph Niederau
MVZ Labor Dortmund Leopoldstraße GbR
Leopoldstraße 10
44147 Dortmund

Tel.: +49 231 860 27 0
Fax: +49 231 860 27 2810
Mail: christoph.niederau@labor-dortmund.de

296

Vitamins 03 - Vitamin C

Analyte	Sam- ple	Unit	Your value	Target value	TV- Type	Lower limit	Upper limit	Deviation	Z-Score	Meets criteria	
Ascorbic acid (Vit. C)	31	mg/l	22.9	29.8	SV	19.1	40.5	-23.2%	-1,69	+	
	32		14.8	19.9	SV	12.7	27.1	-25.6%	-1,92	+	

Analyte	Sam- ple	Your unit	Stated value	Conversion factor	Method	Manu- facturer	Device
Ascorbic acid (Vit. C)	31	mg/l	22.9	1.00	9999	SD	ZY99
	32		14.8	1.00			



Individual summary of results

51396: Dr. Endre Mikus
LabMagister Training and Science Ltd.

Survey of 29 May 2019

296

Vitamins 03 - Vitamin C

Ascorbic acid (Vit. C) (mg/l, N = 63)										
Collective	Sample	Target value	Target range			Participants collective			Rate (%)	
						AVG	CV	Num.	Sam.	total
all methods	31	29.8	19.1	-	40.5	29.8	13.7	63	96.8	95.2
	32	19.9	12.7	-	27.1	19.9	13.3	63	96.8	

Rate of success: 95,2%

